Tech & AI Systems | Jan 2, 2026 | 7 min read | By Bryan White

Soft Exosuits vs. Rigid Frames: A New Era for Rehabilitation Engineering

Public Health * AI * Robotics

Introduction: Redefining the Human-Machine Interface

For half a century, wearable robotics was dominated by rigid anthropomorphic exoskeletons -- external frames of metal and carbon fiber driven by heavy actuators. While rigid frames enable individuals with complete paralysis to stand, they hit a "metabolic wall" in users with residual mobility, including soldiers, stroke survivors, and logistics workers. Because human gait relies on inverted-pendulum mechanics and elastic tendon recoil, adding rigid mass to the distal lower limbs sharply increases leg-swing inertia.[1] Furthermore, simple mechanical hinges cannot match the rolling-and-sliding kinematics of biological joints, creating shear forces that elevate walking energy expenditure.[2]

To resolve this paradox, Professor Conor Walsh and the Harvard Biodesign Lab at the John A. Paulson School of Engineering and Applied Sciences and the Wyss Institute pioneered the Soft Robotic Exosuit.[3] Because ambulatory users already possess a load-bearing skeleton, the exosuit discards the rigid frame and functions as an external layer of muscles and tendons -- routing tensile forces through functional textiles while the user's bones bear compression.[3][4] Mounting motors and batteries near the waist and keeping distal textile wraps under a kilogram makes the "Un-Exoskeleton" mechanically transparent when slack and metabolically advantageous when active.[3][10] This report traces the evolution of soft exosuits from DARPA prototypes to FDA-cleared stroke rehabilitation systems and industrial safety apparel.

Engineering Architecture: Functional Textiles and Bowden Actuation

Transmitting over 200 Newtons of tensile force across compliant human tissue without slippage or pain required an integrated soft-hard engineering stack:[4][7]

* Load Paths & Series Elasticity: High-stiffness nylon webbing and rip-stop weaves route forces between a padded pelvic belt and a high-friction calf/boot anchor.[4][5] The textile and underlying tissue act as a natural Series Elastic Actuator (SEA), absorbing shock loads and storing elastic strain energy.[4]

* Bowden Cable Drives: Replacing inefficient pneumatic actuators burdened by air compressibility and thermodynamic losses,[3] waist-mounted brushless DC motors retract spooled Bowden cables to deliver ~30% of biological joint moments at bandwidths above 20 Hz without distal leg mass.[6][7][8]

* Inertial & Soft Sensing: Thigh- and shank-mounted Inertial Measurement Units (IMUs), load cells, and hyper-elastic liquid-metal strain sensors detect heel strike and toe-off within milliseconds to synchronize cable tension with gait events.[6][9][4]

Human-in-the-Loop Control and Metabolic Economies

Because optimal assistance timing varies by 10-20% of the gait cycle across individuals, the Harvard team introduced Human-in-the-Loop (HIL) optimization, feeding respiratory oxygen consumption (VO2) into machine learning algorithms that personalize control parameters online:[10][12]

| Optimization Algorithm | | Mechanism | | Data Efficiency & Noise Tolerance | | Convergence | | Primary Use Case |

| Gradient Descent | | Follows local slope of cost surface | | Low; sensitive to metabolic noise | | Slow (>45 mins) | | Low-noise benchtop emulators |

| Bayesian Optimization | | Gaussian Process surrogate model | | High; handles sparse, noisy data | | Fast (~21 mins)[10][12] | | Clinical tuning & rapid personalization |

| CMA-ES | | Evolves population covariance matrix | | Medium; excellent outlier robustness | | Medium (~35 mins) | | Complex multi-joint spaces[13] |

Combining bio-inspired torque profiles with HIL tuning shattered the metabolic barrier.[2][11] In a landmark 2017 Science Robotics study of unloaded level walking (1.5 m/s), a tethered multiarticular soft exosuit delivering peak bilateral ankle and indirect hip assistance reduced net metabolic cost by 22.8% +/- 3.17% relative to the powered-off condition.[14][15]

Clinical and Industrial Translation: ReStore and Verve Motion

In post-stroke hemiparesis, patients exhibit drop foot (toe drag during swing) and stance propulsion deficits that force compensatory hip hiking and circumduction.[16] Co-developed with ReWalk Robotics (now Lifeward), the unilateral ReStore exosuit uses anterior and posterior cables to assist both dorsiflexion and plantarflexion.[17] Clinical trials demonstrated a ~20% improvement in propulsive symmetry and faster walking speeds, earning historic FDA 510(k) clearance in June 2019.[16][17][18]

In 2020, Harvard spin-out Verve Motion commercialized the backpack-style SafeLift exosuit for logistics workers.[19] Delivering up to 240 N of lumbar extension force, SafeLift offloads 40% of lifted weight from the spine -- reducing workplace back injuries by 60-85% and raising productivity by 4-8% across 65 million commercial lifts.[20][21]

| Era & Platform | | Core Architecture | | Measured Physiological / Clinical Outcome |

| 2012-2017: DARPA & Multi-Joint Suit | | Waist-mounted Bowden cables + HIL Bayesian tuning | | 22.8% +/- 3.17% net metabolic reduction in unloaded level walking (tethered multiarticular suit).[3][14][15] |

| 2018-2020: ReStore Stroke System | | Unilateral dual-cable plantarflexion/dorsiflexion | | FDA 510(k) cleared (2019); ~20% gain in post-stroke propulsion symmetry.[16][17][18] |

| 2021-2025+: Verve SafeLift & AI | | Lumbar-assist cables & predictive ML actuators | | 240 N lift assist (40% offload); 60-85% fewer back injuries across 65M+

lifts.[20][21][24] |

Conclusion: The Future of Soft Wearable Robotics

Ongoing engineering challenges include mitigating localized skin shear, cuff migration, and thermal burden under foam wraps, alongside replacing Bowden cableboxes with fully soft electro-hydraulic actuators.[4][22][23] As the Harvard Biodesign Lab integrates intent-predicting AI for aging-in-place mobility apparel, soft exosuits continue to dissolve the boundary between clothing and robotics -- augmenting human movement through biomechanical synergy rather than rigid force.[1][24]

Updates... (Nov 24, 2026): Corrected the experimental condition for Quinlivan et al. (2017) from loaded walking to unloaded level walking (1.5 m/s) with a tethered multiarticular soft exosuit; replaced the HTTP 400 FDA query link and timed-out NICE PDF link with the live FDA 510(k) K190337 summary PDF and canonical NICE MIB239 overview page; added missing co-authors and canonical DOI for Kulkarni et al. (2025); and added exact online/print dates for Asbeck et al. (2015) and Strick et al. (2024). (Prior Nov 21, 2026 update verified references against CrossRef/PubMed and added anchor IDs.)

References & Primary Sources

* [1] Wyss Institute (Jan 18, 2017). Soft exosuit economies: understanding the costs of lightening the load. -- https://wyss.harvard.edu/news/soft-exosuit-economies-understanding-the-costs-of-lightening-the-load/

* [2] Harvard Magazine (Dec 9, 2013). Harvard engineers design soft robotic exosuits to aid human movement. -- https://www.harvardmagazine.com/2013/12/wearable-robots

* [3] Harvard John A. Paulson School of Engineering and Applied Sciences (Sep 11, 2014). Walsh to advance wearable robot design. -- https://seas.harvard.edu/news/2014/09/walsh-advance-wearable-robot-design

* [4] Asbeck, A. T., De Rossi, S. M. M., Holt, K. G., & Walsh, C. J. (Mar 16, 2015; print May 2015). A biologically inspired soft exosuit for walking assistance. The International Journal of Robotics Research. -- https://doi.org/10.1177/0278364914562476

* [5] Walsh, C. J., Asbeck, A. T., & De Rossi, S. M. M. (May 31, 2016). US9351900B2 - Soft exosuit for assistance with human motion. Google Patents. -- https://patents.google.com/patent/US9351900B2/en

* [6] Yandell, M. B., Quinlivan, B. T., Popov, D., Walsh, C., & Zelik, K. E. (May 18, 2017). Physical interface dynamics alter how robotic exosuits augment human movement: implications for optimizing wearable assistive devices. Journal of NeuroEngineering and Rehabilitation, 14(1), 40. -- https://doi.org/10.1186/s12984-017-0247-9

* [7] Asbeck, A. T., Dyer, R. J., Larusson, A. F., & Walsh, C. J. (Jun 2013). Biologically-inspired soft exosuit. IEEE 13th International Conference on Rehabilitation Robotics (ICORR). -- https://doi.org/10.1109/icorr.2013.6650455

* [8] Yang, J., Moon, J., Ryu, J., Kim, J., Nam, K., Park, S., Kim, Y., & Lee, G.

(Jan 3, 2023). Design of a Quasi-Direct Drive Actuator with Embedded Pulley for a Compact, Lightweight, and High-Bandwidth Exosuit. Actuators, 12(1), 21. -- https://doi.org/10.3390/act12010021

* [9] Strick, J. A., Farris, R. J., & Sawicki, J. T. (Jan 29, 2024; print Apr 1, 2024). A Novel Gait Event Detection Algorithm Using a Thigh-Worn Inertial Measurement Unit and Joint Angle Information. Journal of Biomechanical Engineering, 146(4), 044502. -- https://doi.org/10.1115/1.4064435

* [10] Ding, Y., Kim, M., Kuindersma, S., & Walsh, C. J. (Feb 28, 2018). Human-in-the-loop optimization of hip assistance with a soft exosuit during walking. Science Robotics, 3(15), eaar5438. -- https://doi.org/10.1126/scirobotics.aar5438

* [11] Ding, Y., Galiana, I., Siviy, C., Panizzolo, F. A., & Walsh, C. (May 2016). IMU-based iterative control for hip extension assistance with a soft exosuit. IEEE International Conference on Robotics and Automation (ICRA). -- https://doi.org/10.1109/icra.2016.7487530

* [12] Kim, M., Ding, Y., Malcolm, P., et al. (Sep 19, 2017). Human-in-the-loop Bayesian optimization of wearable device parameters. PLOS One. -- https://doi.org/10.1371/journal.pone.0184054

* [13] Hansen, N., & Ostermeier, A. (Jun 2001). Completely Derandomized Self-Adaptation in Evolution Strategies. Evolutionary Computation, 9(2), 159-195. -- https://doi.org/10.1162/106365601750190398

* [14] Quinlivan, B. T., Lee, S., Malcolm, P., et al. (Jan 18, 2017). Assistance magnitude versus metabolic cost reductions for a tethered multiarticular soft exosuit. Science Robotics, 2(2), eaah4416. -- https://doi.org/10.1126/scirobotics.aah4416

* [15] Harvard John A. Paulson School of Engineering and Applied Sciences (Jan 18, 2017). Soft exosuit economies: understanding the costs of lightening the load. -- https://seas.harvard.edu/news/2017/01/soft-exosuit-economies-understanding-costs-lightening-load

* [16] National Institute for Health and Care Excellence (NICE) (Dec 15, 2020). ReStore Soft Exo-Suit for gait rehabilitation (Medtech innovation briefing MIB239). -- https://www.nice.org.uk/advice/mib239

* [17] Awad, L. N., Esquenazi, A., Francisco, G. E., et al. (Jun 18, 2020). The ReWalk ReStore soft robotic exosuit: a multi-site clinical trial of the safety, reliability, and feasibility of exosuit-augmented post-stroke gait rehabilitation. Journal of NeuroEngineering and Rehabilitation, 17, 80. -- https://doi.org/10.1186/s12984-020-00702-5

* [18] U.S. Food and Drug Administration (FDA) (Jun 3, 2019). 510(k) Premarket Notification Summary (K190337): ReWalk ReStore Soft Exo-Suit. -- https://www.accessdata.fda.gov/cdrh_docs/pdf19/K190337.pdf

* [19] PR Newswire / Verve Motion (Jul 15, 2021). Wearable Robotics Startup Verve Motion Snags $15M Series A (SafeLift Exosuit Commercial Launch). -- https://www.prnewswire.com/news-releases/wearable-robotics-startup-verve-motion-snags-15m-series-a-301334795.html

* [20] Verve Motion (Dec 15, 2023; updated Apr 18, 2025). Real Stories of Verve Motion's SafeLift Exosuit in the Workplace. --

https://vervemotion.com/customer-results/

* [21] Robotics 24/7 (Jul 19, 2024). Verve Motion releases study on soft exosuits. -- https://www.robotics247.com/article/verve_motion_releases_study_on_soft_exosuits

* [22] Kulkarni, M., Edward, S., Golecki, T., Kaehr, B., & Golecki, H. (Feb 20, 2025). Soft robots built for extreme environments. Soft Science, 5(1). -- https://doi.org/10.20517/ss.2023.51

* [23] Rykaczewski, K. (Aug 23, 2022). Thermophysiological aspects of wearable robotics: Challenges and opportunities. Temperature. -- https://doi.org/10.1080/23328940.2022.2113725

* [24] MassAITC (Jul 2, 2025). 2025 a2 National Symposium Plenary Talk: Conor Walsh, PhD -- ML/AI in Soft Wearable Robots. -- https://massaitc.org/2025/07/02/2025-a2-national-symposium-plenary-talk-conor-walsh-phd-machine-learning-and-ai-methods-for-a